

**Industrie Service
Energie und Umwelt**



**TÜV Rheinland Immissionsschutz und Energiesysteme GmbH
D-51101 Köln**

Met One Instruments, Inc.
Dr. Herbert Schloesser
1600 NW Washington Blvd.

Grants Pass, OR 97526

USA

Mr. Karsten Pletscher
Phone ++49 221 806-2592
Fax ++49 221 806-1349
Mail karsten.pletscher
@de.tuv.com
Cologne, May 06, 2010

- sent in advance as pdf-document by e-mail -

**Status of the suitability test of the ambient air quality measuring system
BAM-1020 of the company Met One Instruments, Inc. for the measured component
PM_{2.5}**

Dear Dr. Schloesser,

herewith we give you an overview on the actual status of the suitability test of your ambient air quality measuring system BAM-1020 for the measured component PM_{2.5}. The mentioned system is a part of the combined UK-Germany PM Equivalence Program, which is a joint test work in order to achieve both British MCERTS and German UBA certification.

The tested systems have passed successfully the test work, consisting of 4 field test sites (Teddington (UK) – Summer, Cologne (GER) – Winter, Bornheim (GER) – Summer, Teddington (UK) - Winter) as well as the laboratory test.

The test report has recently achieved an overall positive assessment at the German technical committee "Test reports". This means, that the particulate monitor system BAM-1020 for the measured component PM_{2.5} can henceforth be characterized as a suitability-tested measuring system. The official announcement in the German "Bundesanzeiger" is to expect during the next weeks.

If there are further questions, don't hesitate to contact us.

Yours Sincerely

Environmental Protection

i. V.

Dr. rer. nat. Peter Wilbring

Dipl.-Ing. Karsten Pletscher

TÜV Rheinland Immissions-
schutz und Energiesysteme
GmbH
Am Grauen Stein
D-51105 Köln

Tel ++49 221 806-2756
Fax ++49 221 806-1349
Mail luft@de.tuv.com
Web www.umwelt-tuv.de

Geschäftsführung
Dr.-Ing. Wolfgang Jockel

Amtsgericht Köln HRB 32190