

Met One Instruments, Inc. Issued US-EPA PM-Coarse Designation  
on BAM-1020 Monitors

Met One Instruments, Inc. is pleased to announce that it has been awarded PM-Coarse designation on paired BAM-1020 monitors. The BAM-1020 PM-Coarse Measurement System consists of interconnected BAM-1020 monitors collocated for making the PM-Coarse measurement. The individual BAM-1020 monitors are designated PM<sub>2.5</sub> (EQPM-0308-170) and PM<sub>10</sub> (EQPM-0798-122) equivalent methods in their own right and may be separated at any time in order to perform these measurements independently.

For further information, please contact Dr. David Gobeli at 541-471-7111 x 102 or [dgobeli@metone.com](mailto:dgobeli@metone.com).